

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121721\
 Data File : VW024068.D
 Acq On : 17 Dec 2021 16:40
 Operator : SY/MD
 Sample : M5133-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBH6

Quant Time: Dec 18 01:47:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Dec 18 01:42:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	83624	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	89442	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	44149	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	19312	4.296	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.000%	
7) Chloroethane-d5	1.565	69	18244	4.864	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	97.200%	
11) 1,1-Dichloroethene-d2	2.105	63	32549	3.709	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.200%	
20) 2-Butanone-d5	3.896	46	47470	54.048	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.100%	
24) Chloroform-d	4.349	84	48805	4.793	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
26) 1,2-Dichloroethane-d4	5.034	65	25714	5.424	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.400%	
32) Benzene-d6	5.050	84	89746	5.066	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.400%	
36) 1,2-Dichloropropane-d6	6.069	67	26472	4.953	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.000%	
41) Toluene-d8	7.317	98	78330	4.911	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.200%	
43) trans-1,3-Dichloroprop...	7.625	79	9720	4.384	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.600%	
46) 2-Hexanone-d5	8.088	63	38231	38.516	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	77.040%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	18403	4.151	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	83.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	31678	5.372	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.400%	
Target Compounds						
					Qvalue	
13) Acetone	2.182	43	9831	13.991	ug/L	87
31) Carbon tetrachloride	4.828	117	86231	7.839	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121721\
Data File : VV024068.D
Acq On : 17 Dec 2021 16:40
Operator : SY/MD
Sample : M5133-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBBH6

Quant Time: Dec 18 01:47:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Dec 18 01:42:59 2021
Response via : Initial Calibration

