

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121818\
 Data File : VV009024.D
 Acq On : 18 Dec 2018 19:43
 Operator : SY/MD
 Sample : J6418-15
 Misc : 4.52G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEZZ3

Quant Time: Dec 19 07:14:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM120618S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 19 06:18:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	251260	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	222597	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	89914	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	63106	24.27	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.08%
7) Chloroethane-d5	1.58	69	60652	31.57	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	126.28%
10) 1,1-Dichloroethene-d2	2.13	63	129186	19.87	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.48%
20) 2-Butanone-d5	3.94	46	40297	42.57	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.14%
24) Chloroform-d	4.40	84	162588	26.92	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	107.68%
26) 1,2-Dichloroethane-d4	5.08	65	96861	29.39	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	117.56%
29) Benzene-d6	5.10	84	326534	26.26	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.04%
33) 1,2-Dichloropropane-d6	6.12	67	99903	28.70	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	114.80%
37) Toluene-d8	7.36	98	294592	25.60	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.40%
38) trans-1,3-Dichloropropene-	7.66	79	38956	25.47	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	101.88%
39) 2-Hexanone-d5	8.14	63	24510	38.44	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.88%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	69255	23.72	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.88%
61) 1,2-Dichlorobenzene-d4	11.67	152	87295	25.87	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.48%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121818\
Data File : VV009024.D
Acq On : 18 Dec 2018 19:43
Operator : SY/MD
Sample : J6418-15
Misc : 4.52G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEZZ3

Quant Time: Dec 19 07:14:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM120618S.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 19 06:18:48 2018
Response via : Initial Calibration

