

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121818\
 Data File : VV009025.D
 Acq On : 18 Dec 2018 20:07
 Operator : SY/MD
 Sample : J6418-16
 Misc : 4.81G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEZZ5

Quant Time: Dec 19 07:16:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM120618S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 19 06:18:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	262418	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	234169	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	95723	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	62170	22.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.56%
7) Chloroethane-d5	1.58	69	59791	29.80	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	119.20%
10) 1,1-Dichloroethene-d2	2.13	63	118305	17.42	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.68%
20) 2-Butanone-d5	3.94	46	44184	44.69	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.38%
24) Chloroform-d	4.40	84	165611	26.25	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.00%
26) 1,2-Dichloroethane-d4	5.08	65	102636	29.81	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	119.24%
29) Benzene-d6	5.10	84	330530	25.27	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.08%
33) 1,2-Dichloropropane-d6	6.12	67	101409	27.70	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	110.80%
37) Toluene-d8	7.36	98	295097	24.38	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.52%
38) trans-1,3-Dichloropropene-	7.67	79	41448	25.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	103.04%
39) 2-Hexanone-d5	8.14	63	24674	36.79	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.58%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	76100	24.78	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.12%
61) 1,2-Dichlorobenzene-d4	11.68	152	91203	25.39	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.56%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121818\
Data File : VV009025.D
Acq On : 18 Dec 2018 20:07
Operator : SY/MD
Sample : J6418-16
Misc : 4.81G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEZZ5

Quant Time: Dec 19 07:16:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM120618S.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 19 06:18:48 2018
Response via : Initial Calibration

