

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121819\
 Data File : VV014115.D
 Acq On : 18 Dec 2019 12:35
 Operator : SY/MD
 Sample : K6314-12
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFF47

Quant Time: Dec 19 07:48:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 19 07:06:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	719404	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	704247	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	318433	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	215924	51.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.80%
7) Chloroethane-d5	1.58	69	204239	47.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.06%
11) 1,1-Dichloroethene-d2	2.13	63	301692	32.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.58%
21) 2-Butanone-d5	3.92	46	296689	101.31	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.31%
24) Chloroform-d	4.40	84	425022	48.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.28%
26) 1,2-Dichloroethane-d4	5.08	65	271339	49.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.02%
32) Benzene-d6	5.10	84	835184	45.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.54%
36) 1,2-Dichloropropane-d6	6.11	67	266854	48.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.44%
41) Toluene-d8	7.36	98	707380	42.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.04%
43) trans-1,3-Dichloropropene-	7.66	79	110508	41.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.92%
47) 2-Hexanone-d5	8.13	63	190651	93.21	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.21%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	356085	44.89	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.78%
64) 1,2-Dichlorobenzene-d4	11.67	152	296905	47.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.94%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121819\
Data File : VV014115.D
Acq On : 18 Dec 2019 12:35
Operator : SY/MD
Sample : K6314-12
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFF47

Quant Time: Dec 19 07:48:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 19 07:06:05 2019
Response via : Initial Calibration

