

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121820\
 Data File : VV019709.D
 Acq On : 18 Dec 2020 18:56
 Operator : SY/MD
 Sample : L5101-05DL 400X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0FX1DL

Quant Time: Dec 19 08:10:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 19 05:26:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	591244	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	555712	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	256870	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	176682	41.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.34%
7) Chloroethane-d5	1.57	69	153953	46.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.08%
11) 1,1-Dichloroethene-d2	2.11	63	266436	33.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.72%
21) 2-Butanone-d5	3.90	46	309428	134.28	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	134.28%#
24) Chloroform-d	4.35	84	351919	46.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.12%
26) 1,2-Dichloroethane-d4	5.04	65	248876	49.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.68%
32) Benzene-d6	5.05	84	704821	47.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.10%
36) 1,2-Dichloropropane-d6	6.07	67	230161	49.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.14%
41) Toluene-d8	7.32	98	609084	45.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.28%
43) trans-1,3-Dichloropropene-	7.62	79	109346	41.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.82%
47)						

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121820\
Data File : VV019709.D
Acq On : 18 Dec 2020 18:56
Operator : SY/MD
Sample : L5101-05DL 400X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0FX1DL

Quant Time: Dec 19 08:10:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 19 05:26:51 2020
Response via : Initial Calibration

