

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121820\
 Data File : VV019712.D
 Acq On : 18 Dec 2020 21:04
 Operator : SY/MD
 Sample : VV1218WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK73

Quant Time: Dec 19 08:13:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 19 07:00:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	597213	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	559589	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	267237	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	188448	44.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.00%
7) Chloroethane-d5	1.57	69	158351	46.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.76%
11) 1,1-Dichloroethene-d2	2.11	63	271095	34.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.22%
21) 2-Butanone-d5	3.89	46	316785	136.10	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	136.10%#
24) Chloroform-d	4.35	84	354059	45.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.74%
26) 1,2-Dichloroethane-d4	5.04	65	245460	48.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.32%
32) Benzene-d6	5.05	84	705839	47.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.58%
36) 1,2-Dichloropropane-d6	6.07	67	230784	49.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.72%
41) Toluene-d8	7.32	98	611315	45.49	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.98%
43) trans-1,3-Dichloropropene-	7.62	79	107945	40.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.18%
47) 2-Hexanone-d5	8.09	63	209834	114.36	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	114.36%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	298517	47.21	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.42%
64) 1,2-Dichlorobenzene-d4	11.63	152	229211	45.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.76%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121820\
Data File : VV019712.D
Acq On : 18 Dec 2020 21:04
Operator : SY/MD
Sample : VV1218WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK73

Quant Time: Dec 19 08:13:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 19 07:00:18 2020
Response via : Initial Calibration

