

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121919\
 Data File : VV014129.D
 Acq On : 19 Dec 2019 15:00
 Operator : SY/MD
 Sample : K6286-01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFDX7

Quant Time: Dec 20 07:34:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 20 07:12:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	367834	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	346670	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	149512	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94986	4.17	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.40%
7) Chloroethane-d5	1.58	69	97211	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.60%
11) 1,1-Dichloroethene-d2	2.13	63	123697	3.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.60%
20) 2-Butanone-d5	3.95	46	253990	53.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.34%
24) Chloroform-d	4.39	84	219874	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.07	65	108344	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	5.09	84	373534	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.60%
36) 1,2-Dichloropropane-d6	6.11	67	127622	4.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.80%
41) Toluene-d8	7.35	98	264593	3.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.00%#
43) trans-1,3-Dichloropropene-	7.66	79	40122	3.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.40%
46) 2-Hexanone-d5	8.13	63	157908	41.26	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.52%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	95302	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	126984	4.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.20%

Target Compounds					Ovalue
13) Acetone	2.22	43	9672	2.922 ug/L	89
42) Toluene	7.43	91	18861	0.207 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121919\
Data File : VV014129.D
Acq On : 19 Dec 2019 15:00
Operator : SY/MD
Sample : K6286-01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFDX7

Quant Time: Dec 20 07:34:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 20 07:12:45 2019
Response via : Initial Calibration

