

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121919\
 Data File : VV014139.D
 Acq On : 19 Dec 2019 18:59
 Operator : SY/MD
 Sample : K6286-11
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFDY7

Quant Time: Dec 20 08:13:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 20 07:12:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	335653	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	340539	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	152107	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	88422	4.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.20%
7) Chloroethane-d5	1.58	69	88692	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.60%
11) 1,1-Dichloroethene-d2	2.13	63	116245	3.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.40%
20) 2-Butanone-d5	3.93	46	297294	68.84	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	137.68%#
24) Chloroform-d	4.40	84	210877	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.08	65	105078	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
32) Benzene-d6	5.10	84	352619	4.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.20%
36) 1,2-Dichloropropane-d6	6.11	67	119353	4.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.20%
41) Toluene-d8	7.36	98	262477	3.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.60%#
43) trans-1,3-Dichloropropene-	7.66	79	40037	3.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.60%
46) 2-Hexanone-d5	8.13	63	196774	52.35	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.70%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	95849	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	123014	4.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.60%

Target Compounds

					Ovalue
13) Acetone	2.19	43	8754	2.899 ug/L #	56
25) Chloroform	4.42	83	41039	1.064 ug/L	98
34) Trichloroethene	5.96	95	5369	0.233 ug/L	89
38) Bromodichloromethane	6.55	83	5817	0.205 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121919\
Data File : VV014139.D
Acq On : 19 Dec 2019 18:59
Operator : SY/MD
Sample : K6286-11
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFDY7

Quant Time: Dec 20 08:13:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 20 07:12:45 2019
Response via : Initial Calibration

