

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121921\
 Data File : VW024094.D
 Acq On : 19 Dec 2021 17:27
 Operator : SY/MD
 Sample : M5134-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBK7

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/20/2021
 Supervised By : Mahesh Dadoda 12/22/2021

Quant Time: Dec 20 01:38:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 20 01:35:07 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	93718	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	98376	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60308	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	26916	5.342	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 106.800%		
7) Chloroethane-d5	1.568	69	24163	5.748	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 115.000%		
11) 1,1-Dichloroethene-d2	2.108	63	43766	4.450	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 89.000%		
20) 2-Butanone-d5	3.915	46	46394	47.133	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 94.260%		
24) Chloroform-d	4.349	84	68268	5.982	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 119.600%		
26) 1,2-Dichloroethane-d4	5.034	65	32122	6.045	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 121.000%		
32) Benzene-d6	5.050	84	127262	6.531	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 130.600%#		
36) 1,2-Dichloropropane-d6	6.069	67	37267	6.339	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 126.800%		
41) Toluene-d8	7.317	98	118248	6.741	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 134.800%#		
43) trans-1,3-Dichloroprop...	7.625	79	13477	5.526	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 110.600%		
46) 2-Hexanone-d5	8.092	63	48869	44.763	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 89.520%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	25764	5.283	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 105.600%		
66) 1,2-Dichlorobenzene-d4	11.625	152	51788	6.429	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 128.600%#		
Target Compounds						Qvalue
13) Acetone	2.220	43	15882m	20.168	ug/L	
14) Carbon disulfide	2.298	76	4044	0.222	ug/L	99
15) Methyl Acetate	2.442	43	23803	12.470	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
Data File : VV024094.D
Acq On : 19 Dec 2021 17:27
Operator : SY/MD
Sample : M5134-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBBK7

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/20/2021
Supervised By : Mahesh Dadoda 12/22/2021

Quant Time: Dec 20 01:38:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Dec 20 01:35:07 2021
Response via : Initial Calibration

