




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
 Data File : VV024108.D
 Acq On : 19 Dec 2021 23:26
 Operator : SY/MD
 Sample : M5148-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBR9

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 12/20/2021
 Supervised By :Mahesh Dadoda 12/22/2021

Quant Time: Dec 20 04:19:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 20 04:04:22 2021
 Response via : Initial Calibration



TIC: VV024108.D\data.ms

(13) Acetone (T)

2.211min (+ 0.006) 6.76 ug/L

response 4695

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	20.70	25.22
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
 Data File : VV024108.D
 Acq On : 19 Dec 2021 23:26
 Operator : SY/MD
 Sample : M5148-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBR9

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 12/20/2021
 Supervised By :Mahesh Dadoda 12/22/2021

Quant Time: Dec 20 04:19:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 20 04:04:22 2021
 Response via : Initial Calibration



TIC: VV024108.D\data.ms

(13) Acetone (T)

2.211min (+ 0.006) 9.04 ug/L m

response 6279

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	20.70	18.86
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
 Data File : VV024108.D
 Acq On : 19 Dec 2021 23:26
 Operator : SY/MD
 Sample : M5148-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBR9

Manual IntegrationsAPPROVED

Reviewed By : John Carlone 12/20/2021
 Supervised By : Mahesh Dadoda 12/22/2021

Quant Time: Dec 20 04:19:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 20 04:04:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	82686	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	85309	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	41272	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	17328	3.898	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	78.000%		
7) Chloroethane-d5	1.567	69	16533	4.458	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	89.200%		
11) 1,1-Dichloroethene-d2	2.108	63	27073	3.120	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	62.400%		
20) 2-Butanone-d5	3.928	46	33830	38.955	ug/L	0.02
Spiked Amount 50.000	Range 40 - 130		Recovery =	77.900%		
24) Chloroform-d	4.352	84	45398	4.509	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	90.200%		
26) 1,2-Dichloroethane-d4	5.037	65	23490	5.011	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	100.200%		
32) Benzene-d6	5.050	84	81512	4.824	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	96.400%		
36) 1,2-Dichloropropane-d6	6.069	67	24223	4.752	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	95.000%		
41) Toluene-d8	7.316	98	73693	4.844	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	96.800%		
43) trans-1,3-Dichloroprop...	7.625	79	8605	4.069	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	81.400%		
46) 2-Hexanone-d5	8.091	63	31319	33.082	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	66.160%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	17214	4.071	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	81.400%		
66) 1,2-Dichlorobenzene-d4	11.622	152	30154	5.470	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	109.400%		
Target Compounds						
13) Acetone	2.211	43	6279m	9.037	ug/L	
22) cis-1,2-Dichloroethene	3.921	96	2295	0.353	ug/L	97
34) Trichloroethene	5.915	95	33272	4.888	ug/L	98
42) Toluene	7.397	91	4773	0.176	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
Data File : VV024108.D
Acq On : 19 Dec 2021 23:26
Operator : SY/MD
Sample : M5148-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBBR9

Manual IntegrationsAPPROVED

Quant Time: Dec 20 04:19:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Dec 20 04:04:22 2021
Response via : Initial Calibration

Reviewed By :John Carlone 12/20/2021
Supervised By :Mahesh Dadoda 12/22/2021

