




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
 Data File : VV024121.D
 Acq On : 20 Dec 2021 04:35
 Operator : SY/MD
 Sample : M5148-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBT0

Manual IntegrationsAPPROVED

Quant Time: Dec 20 05:38:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 20 04:04:22 2021
 Response via : Initial Calibration

Reviewed By :John Carlone 12/20/2021
 Supervised By :Mahesh Dadoda 12/22/2021



TIC: VV024121.D\data.ms

(13) Acetone (T)

2.211min (+ 0.006) 8.02 ug/L

response 5817

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	20.70	16.50
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
 Data File : VV024121.D
 Acq On : 20 Dec 2021 04:35
 Operator : SY/MD
 Sample : M5148-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBT0

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 12/20/2021
 Supervised By :Mahesh Dadoda 12/22/2021

Quant Time: Dec 20 05:38:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 20 04:04:22 2021
 Response via : Initial Calibration



TIC: VV024121.D\data.ms

(13) Acetone (T)

2.211min (+ 0.006) 10.76 ug/L m

response 7804

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	20.70	12.30
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
 Data File : VV024121.D
 Acq On : 20 Dec 2021 04:35
 Operator : SY/MD
 Sample : M5148-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBT0

Manual IntegrationsAPPROVED

Quant Time: Dec 20 05:38:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 20 04:04:22 2021
 Response via : Initial Calibration

Reviewed By : John Carlone 12/20/2021
 Supervised By : Mahesh Dadoda 12/22/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	86301	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	94372	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	46088	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	19525	4.208	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	84.200%	
7) Chloroethane-d5	1.568	69	18776	4.851	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	97.000%	
11) 1,1-Dichloroethene-d2	2.108	63	31661	3.496	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	70.000%	
20) 2-Butanone-d5	3.921	46	30073	33.178	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	66.360%	
24) Chloroform-d	4.349	84	52253	4.972	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.400%	
26) 1,2-Dichloroethane-d4	5.037	65	25310	5.173	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.400%	
32) Benzene-d6	5.053	84	91840	4.913	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.200%	
36) 1,2-Dichloropropane-d6	6.069	67	28685	5.087	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.800%	
41) Toluene-d8	7.317	98	78285	4.652	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.000%	
43) trans-1,3-Dichloroprop...	7.625	79	9143	3.908	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.200%	
46) 2-Hexanone-d5	8.091	63	29536	28.202	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	56.400%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	19634	4.197	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	84.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	32394	5.262	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.200%	
Target Compounds						
					Qvalue	
13) Acetone	2.211	43	7804m	10.762	ug/L	
14) Carbon disulfide	2.294	76	878	0.052	ug/L #	80
30) Cyclohexane	4.680	56	9148	0.934	ug/L #	91
34) Trichloroethene	5.911	95	353364	46.925	ug/L	99
47) Tetrachloroethene	7.973	164	1838	0.277	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121921\
Data File : VV024121.D
Acq On : 20 Dec 2021 04:35
Operator : SY/MD
Sample : M5148-21
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBBT0

Manual IntegrationsAPPROVED

Quant Time: Dec 20 05:38:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Dec 20 04:04:22 2021
Response via : Initial Calibration

Reviewed By :John Carlone 12/20/2021
Supervised By :Mahesh Dadoda 12/22/2021

