

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121922\
 Data File : VW029734.D
 Acq On : 19 Dec 2022 20:07
 Operator : SY/MD
 Sample : N6093-15 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGSN4

Quant Time: Dec 20 00:55:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 20 00:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	453399	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	424271	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	221873	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	121402	43.726	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.460%
7) Chloroethane-d5	1.564	69	103974	49.111	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.220%
11) 1,1-Dichloroethene-d2	2.105	63	164529	37.993	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.980%
21) 2-Butanone-d5	3.873	46	168806	99.667	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.670%
24) Chloroform-d	4.339	84	235018	44.982	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.960%
26) 1,2-Dichloroethane-d4	5.024	65	141538	47.432	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.860%
32) Benzene-d6	5.043	84	501902	44.770	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.540%
36) 1,2-Dichloropropane-d6	6.063	67	159874	46.688	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.380%
41) Toluene-d8	7.310	98	436345	41.309	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.620%
43) trans-1,3-Dichloroprop...	7.612	79	65039	37.613	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.220%
47) 2-Hexanone-d5	8.085	63	135822	92.884	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.880%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	225046	45.990	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.980%
66) 1,2-Dichlorobenzene-d4	11.612	152	185144	43.836	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121922\
Data File : VV029734.D
Acq On : 19 Dec 2022 20:07
Operator : SY/MD
Sample : N6093-15 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGSN4

Quant Time: Dec 20 00:55:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 20 00:50:16 2022
Response via : Initial Calibration

