

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121922\
 Data File : VW029709.D
 Acq On : 19 Dec 2022 10:14
 Operator : SY/MD
 Sample : VW1219WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK311

Quant Time: Dec 20 00:51:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 20 00:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.603	114	509541	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.841	117	470379	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	249532	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	151007	48.396	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	96.800%	
7) Chloroethane-d5	1.561	69	124772	52.442	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	104.880%	
11) 1,1-Dichloroethene-d2	2.101	63	198432	40.773	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	81.540%	
21) 2-Butanone-d5	3.870	46	166802	87.633	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	87.630%	
24) Chloroform-d	4.333	84	261632	44.558	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.120%	
26) 1,2-Dichloroethane-d4	5.018	65	155307	46.312	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.620%	
32) Benzene-d6	5.037	84	563184	45.312	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.620%	
36) 1,2-Dichloropropane-d6	6.056	67	173522	45.706	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	91.420%	
41) Toluene-d8	7.304	98	509873	43.538	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.080%	
43) trans-1,3-Dichloroprop...	7.609	79	72858	38.005	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	76.000%	
47) 2-Hexanone-d5	8.082	63	135037	83.295	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	83.300%	
56) 1,1,2,2-Tetrachloroeth...	10.204	84	229617	42.324	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	84.640%	
66) 1,2-Dichlorobenzene-d4	11.612	152	207562	43.696	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121922\
Data File : VV029709.D
Acq On : 19 Dec 2022 10:14
Operator : SY/MD
Sample : VV1219WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK311

Quant Time: Dec 20 00:51:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 20 00:50:16 2022
Response via : Initial Calibration

