

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121922\
 Data File : VW029711.D
 Acq On : 19 Dec 2022 11:08
 Operator : SY/MD
 Sample : PB149723TB 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VLEB723

Quant Time: Dec 20 00:51:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 20 00:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	492518	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	459797	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	241447	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	145600	48.276	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	96.560%	
7) Chloroethane-d5	1.564	69	116832	50.802	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	101.600%	
11) 1,1-Dichloroethene-d2	2.105	63	185526	39.439	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	78.880%	
21) 2-Butanone-d5	3.873	46	147072	79.938	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	79.940%	
24) Chloroform-d	4.339	84	251446	44.304	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.600%	
26) 1,2-Dichloroethane-d4	5.024	65	147944	45.641	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.280%	
32) Benzene-d6	5.043	84	544088	44.783	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.560%	
36) 1,2-Dichloropropane-d6	6.063	67	169368	45.639	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	91.280%	
41) Toluene-d8	7.307	98	495831	43.314	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.620%	
43) trans-1,3-Dichloroprop...	7.612	79	71332	38.065	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	76.120%	
47) 2-Hexanone-d5	8.082	63	129182	81.517	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	81.520%	
56) 1,1,2,2-Tetrachloroeth...	10.204	84	224398	42.314	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	84.620%	
66) 1,2-Dichlorobenzene-d4	11.612	152	198869	43.268	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.540%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121922\
Data File : VV029711.D
Acq On : 19 Dec 2022 11:08
Operator : SY/MD
Sample : PB149723TB 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VLEB723

Quant Time: Dec 20 00:51:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 20 00:50:16 2022
Response via : Initial Calibration

