

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121922\
 Data File : VW029717.D
 Acq On : 19 Dec 2022 13:29
 Operator : SY/MD
 Sample : N6075-07 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGSM2

Quant Time: Dec 20 00:52:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 20 00:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	534359	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	505753	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	262644	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	143502	43.855	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	87.720%	
7) Chloroethane-d5	1.555	69	90213	36.155	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	72.320%	
11) 1,1-Dichloroethene-d2	2.101	63	196426	38.487	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	76.980%	
21) 2-Butanone-d5	3.873	46	161655	80.984	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	80.980%	
24) Chloroform-d	4.336	84	262992	42.710	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	85.420%	
26) 1,2-Dichloroethane-d4	5.021	65	155274	44.152	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.300%	
32) Benzene-d6	5.040	84	578186	43.265	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	86.540%	
36) 1,2-Dichloropropane-d6	6.059	67	181439	44.449	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	88.900%	
41) Toluene-d8	7.307	98	520347	41.325	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	82.640%	
43) trans-1,3-Dichloroprop...	7.612	79	75050	36.410	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	72.820%	
47) 2-Hexanone-d5	8.085	63	141926	81.421	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	81.420%	
56) 1,1,2,2-Tetrachloroeth...	10.204	84	242322	41.542	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	83.080%	
66) 1,2-Dichlorobenzene-d4	11.612	152	199038	39.810	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	79.620%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121922\
Data File : VV029717.D
Acq On : 19 Dec 2022 13:29
Operator : SY/MD
Sample : N6075-07 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGSM2

Quant Time: Dec 20 00:52:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 20 00:50:16 2022
Response via : Initial Calibration

