

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121922\
 Data File : VW029714.D
 Acq On : 19 Dec 2022 12:18
 Operator : SY/MD
 Sample : N6075-03 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGSM4

Quant Time: Dec 20 00:52:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 20 00:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	489712	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	458336	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	238924	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	179215	59.762	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 119.520%		
7) Chloroethane-d5	1.561	69	147208	64.377	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 128.760%		
11) 1,1-Dichloroethene-d2	2.105	63	238854	51.067	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 102.140%		
21) 2-Butanone-d5	3.873	46	207297	113.318	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 113.320%		
24) Chloroform-d	4.339	84	319849	56.679	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 113.360%		
26) 1,2-Dichloroethane-d4	5.024	65	190011	58.955	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 117.900%		
32) Benzene-d6	5.044	84	711324	58.735	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 117.460%		
36) 1,2-Dichloropropane-d6	6.060	67	220306	59.554	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 119.100%		
41) Toluene-d8	7.307	98	642764	56.328	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 112.660%		
43) trans-1,3-Dichloroprop...	7.613	79	91258	48.853	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 97.700%		
47) 2-Hexanone-d5	8.082	63	175764	111.266	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 111.270%		
56) 1,1,2,2-Tetrachloroeth...	10.204	84	293812	55.580	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 111.160%		
66) 1,2-Dichlorobenzene-d4	11.612	152	261179	57.425	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 114.840%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121922\
Data File : VV029714.D
Acq On : 19 Dec 2022 12:18
Operator : SY/MD
Sample : N6075-03 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGSM4

Quant Time: Dec 20 00:52:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 20 00:50:16 2022
Response via : Initial Calibration

