

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121922\
 Data File : VW029719.D
 Acq On : 19 Dec 2022 14:16
 Operator : SY/MD
 Sample : N6075-09 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGSM5

Quant Time: Dec 20 00:53:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 20 00:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	477889	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	443973	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	227798	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	137870	47.113	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.220%
7) Chloroethane-d5	1.555	69	87365	39.151	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.300%
11) 1,1-Dichloroethene-d2	2.102	63	183828	40.274	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.540%
21) 2-Butanone-d5	3.876	46	160673	90.004	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.000%
24) Chloroform-d	4.339	84	248663	45.155	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.300%
26) 1,2-Dichloroethane-d4	5.021	65	150725	47.922	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.840%
32) Benzene-d6	5.040	84	542750	46.265	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.540%
36) 1,2-Dichloropropane-d6	6.063	67	167970	46.875	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.760%
41) Toluene-d8	7.307	98	487608	44.114	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.220%
43) trans-1,3-Dichloroprop...	7.613	79	71290	39.398	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.800%
47) 2-Hexanone-d5	8.085	63	136567	89.249	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.250%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	233097	45.521	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.040%
66) 1,2-Dichlorobenzene-d4	11.612	152	195095	44.990	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.980%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121922\
Data File : VV029719.D
Acq On : 19 Dec 2022 14:16
Operator : SY/MD
Sample : N6075-09 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGSM5

Quant Time: Dec 20 00:53:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 20 00:50:16 2022
Response via : Initial Calibration

