

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121922\
 Data File : VW029723.D
 Acq On : 19 Dec 2022 15:50
 Operator : SY/MD
 Sample : N6093-03 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGSN8

Quant Time: Dec 20 00:53:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 20 00:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	478803	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	443656	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	232713	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	136360	46.508	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.020%
7) Chloroethane-d5	1.564	69	116121	51.939	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.880%
11) 1,1-Dichloroethene-d2	2.105	63	183871	40.207	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.420%
21) 2-Butanone-d5	3.873	46	175792	98.285	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.290%
24) Chloroform-d	4.339	84	255198	46.253	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.500%
26) 1,2-Dichloroethane-d4	5.024	65	153096	48.583	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.160%
32) Benzene-d6	5.043	84	547281	46.685	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.360%
36) 1,2-Dichloropropane-d6	6.063	67	172478	48.167	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.340%
41) Toluene-d8	7.307	98	491892	44.533	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.060%
43) trans-1,3-Dichloroprop...	7.612	79	70931	39.228	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.460%
47) 2-Hexanone-d5	8.082	63	144209	94.311	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.310%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	236272	46.174	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.340%
66) 1,2-Dichlorobenzene-d4	11.612	152	194245	43.848	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121922\
Data File : VV029723.D
Acq On : 19 Dec 2022 15:50
Operator : SY/MD
Sample : N6093-03 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGSN8

Quant Time: Dec 20 00:53:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 20 00:50:16 2022
Response via : Initial Calibration

