

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122022\
 Data File : VW029751.D
 Acq On : 20 Dec 2022 15:14
 Operator : SY/MD
 Sample : N6130-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Dec 21 00:18:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 21 00:15:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	496534	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.841	117	459264	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	239797	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	133361	43.861	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.720%
7) Chloroethane-d5	1.565	69	118054	50.918	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.840%
11) 1,1-Dichloroethene-d2	2.105	63	184551	38.915	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.820%
21) 2-Butanone-d5	3.873	46	167896	90.518	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.520%
24) Chloroform-d	4.336	84	258030	45.096	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.021	65	158791	48.591	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.180%
32) Benzene-d6	5.040	84	561199	46.245	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.500%
36) 1,2-Dichloropropane-d6	6.060	67	177835	47.976	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.960%
41) Toluene-d8	7.307	98	494848	43.278	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.560%
43) trans-1,3-Dichloroprop...	7.613	79	76561	40.903	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.800%
47) 2-Hexanone-d5	8.082	63	140483	88.752	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.750%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	245550	46.357	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.720%
66) 1,2-Dichlorobenzene-d4	11.612	152	206956	45.337	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122022\
Data File : VV029751.D
Acq On : 20 Dec 2022 15:14
Operator : SY/MD
Sample : N6130-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Dec 21 00:18:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 21 00:15:13 2022
Response via : Initial Calibration

