

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122022\
 Data File : VW029741.D
 Acq On : 20 Dec 2022 11:15
 Operator : SY/MD
 Sample : N6075-17 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGSP3

Quant Time: Dec 21 00:16:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 21 00:15:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	474572	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	440368	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	216793	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	127794	43.975	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.940%
7) Chloroethane-d5	1.565	69	112432	50.737	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.480%
11) 1,1-Dichloroethene-d2	2.105	63	175350	38.686	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.380%
21) 2-Butanone-d5	3.876	46	179766	101.403	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.400%
24) Chloroform-d	4.339	84	254060	46.457	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.920%
26) 1,2-Dichloroethane-d4	5.024	65	152191	48.727	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.460%
32) Benzene-d6	5.043	84	537746	46.214	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.420%
36) 1,2-Dichloropropane-d6	6.063	67	169708	47.748	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.500%
41) Toluene-d8	7.307	98	474879	43.314	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.620%
43) trans-1,3-Dichloroprop...	7.612	79	70246	39.139	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.280%
47) 2-Hexanone-d5	8.085	63	145188	95.660	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.660%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	237369	46.735	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.480%
66) 1,2-Dichlorobenzene-d4	11.612	152	185579	44.968	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122022\
Data File : VV029741.D
Acq On : 20 Dec 2022 11:15
Operator : SY/MD
Sample : N6075-17 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGSP3

Quant Time: Dec 21 00:16:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 21 00:15:13 2022
Response via : Initial Calibration

