

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122023\
 Data File : VW033500.D
 Acq On : 20 Dec 2023 10:42
 Operator : SY/MD
 Sample : 05912-21
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Dec 21 01:45:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM121523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 21 01:43:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	258938	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	266453	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.184	152	111840	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	104594	52.453	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 104.900%		
7) Chloroethane-d5	1.532	69	91813	51.307	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 102.620%		
11) 1,1-Dichloroethene-d2	2.059	65	46099	46.441	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 92.880%		
21) 2-Butanone-d5	3.783	46	110657	110.108	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 110.110%		
24) Chloroform-d	4.246	84	184213	46.702	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 93.400%		
26) 1,2-Dichloroethane-d4	4.940	65	124561	51.406	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.820%		
32) Benzene-d6	4.960	84	356345	49.655	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 99.300%		
36) 1,2-Dichloropropane-d6	5.989	67	114962	52.782	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 105.560%		
41) Toluene-d8	7.243	98	295433	44.991	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 89.980%		
43) trans-1,3-Dichloroprop...	7.554	79	47356	44.364	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 88.720%		
47) 2-Hexanone-d5	8.024	63	67146	91.657	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 91.660%		
56) 1,1,2,2-Tetrachloroeth...	10.152	84	146682	46.857	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 93.720%		
66) 1,2-Dichlorobenzene-d4	11.561	152	110705	52.675	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 105.360%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122023\
Data File : VV033500.D
Acq On : 20 Dec 2023 10:42
Operator : SY/MD
Sample : 05912-21
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Dec 21 01:45:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM121523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 21 01:43:54 2023
Response via : Initial Calibration

