

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122023\
 Data File : VW033506.D
 Acq On : 20 Dec 2023 13:00
 Operator : SY/MD
 Sample : 05850-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0QM0

Quant Time: Dec 21 01:46:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM121523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 21 01:43:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	263039	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	272596	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	114922	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	101582	50.148	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.300%
7) Chloroethane-d5	1.532	69	90417	49.739	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.480%
11) 1,1-Dichloroethene-d2	2.059	65	45241	44.866	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.740%
21) 2-Butanone-d5	3.783	46	94520	92.585	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.590%
24) Chloroform-d	4.249	84	186748	46.607	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.220%
26) 1,2-Dichloroethane-d4	4.940	65	122493	49.764	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.520%
32) Benzene-d6	4.960	84	359105	48.912	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.820%
36) 1,2-Dichloropropane-d6	5.989	67	115465	51.819	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.640%
41) Toluene-d8	7.243	98	304217	45.284	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.560%
43) trans-1,3-Dichloroprop...	7.551	79	48103	44.049	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.100%
47) 2-Hexanone-d5	8.024	63	63241	84.381	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.380%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	144283	45.052	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.100%
66) 1,2-Dichlorobenzene-d4	11.561	152	113468	52.542	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.080%
Target Compounds						
					Qvalue	
51) Chlorobenzene	8.818	112	5201	0.937	ug/L	92
65) 1,4-Dichlorobenzene	11.207	146	3646	0.999	ug/L	90
67) 1,2-Dichlorobenzene	11.583	146	9697	2.747	ug/L	96
70) 1,2,4-trichlorobenzene	13.201	180	4050	1.975	ug/L	96
72) 1,2,3-Trichlorobenzene	13.686	180	2265	1.129	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122023\
Data File : VV033506.D
Acq On : 20 Dec 2023 13:00
Operator : SY/MD
Sample : 05850-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0QM0

Quant Time: Dec 21 01:46:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM121523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 21 01:43:54 2023
Response via : Initial Calibration

