

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122023\
 Data File : VW033507.D
 Acq On : 20 Dec 2023 13:23
 Operator : SY/MD
 Sample : 05850-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0QM1

Quant Time: Dec 21 01:46:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM121523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 21 01:43:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	262766	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	269753	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	113041	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	101984	50.399	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.800%		
7) Chloroethane-d5	1.532	69	92275	50.814	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 101.620%		
11) 1,1-Dichloroethene-d2	2.060	65	45843	45.510	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 91.020%		
21) 2-Butanone-d5	3.783	46	102056	100.071	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 100.070%		
24) Chloroform-d	4.249	84	189231	47.275	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 94.560%		
26) 1,2-Dichloroethane-d4	4.941	65	124095	50.467	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.940%		
32) Benzene-d6	4.960	84	365809	50.350	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.700%		
36) 1,2-Dichloropropane-d6	5.989	67	116919	53.024	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 106.040%		
41) Toluene-d8	7.243	98	302466	45.498	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 91.000%		
43) trans-1,3-Dichloroprop...	7.555	79	49215	45.542	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 91.080%		
47) 2-Hexanone-d5	8.024	63	65152	87.847	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 87.850%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	145462	45.899	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 91.800%		
66) 1,2-Dichlorobenzene-d4	11.561	152	113082	53.234	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.460%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122023\
Data File : VV033507.D
Acq On : 20 Dec 2023 13:23
Operator : SY/MD
Sample : 05850-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0QM1

Quant Time: Dec 21 01:46:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM121523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 21 01:43:54 2023
Response via : Initial Calibration

