

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122023\
 Data File : VW033519.D
 Acq On : 20 Dec 2023 18:04
 Operator : SY/MD
 Sample : 05927-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH3P6

Quant Time: Dec 21 01:49:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM121523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 21 01:43:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	244696	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	261552	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	104898	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	95568	50.716	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.440%
7) Chloroethane-d5	1.532	69	88291	52.210	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.420%
11) 1,1-Dichloroethene-d2	2.060	65	45129	48.110	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.220%
21) 2-Butanone-d5	3.783	46	84825	89.317	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.320%
24) Chloroform-d	4.249	84	182046	48.839	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.680%
26) 1,2-Dichloroethane-d4	4.940	65	117682	51.394	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.780%
32) Benzene-d6	4.960	84	337598	47.924	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.840%
36) 1,2-Dichloropropane-d6	5.985	67	108119	50.571	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.140%
41) Toluene-d8	7.243	98	271913	42.185	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.360%
43) trans-1,3-Dichloroprop...	7.551	79	45126	43.067	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.140%
47) 2-Hexanone-d5	8.024	63	56629	78.749	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.750%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	141575	46.073	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.140%
66) 1,2-Dichlorobenzene-d4	11.561	152	102834	52.168	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.340%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122023\
Data File : VV033519.D
Acq On : 20 Dec 2023 18:04
Operator : SY/MD
Sample : 05927-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH3P6

Quant Time: Dec 21 01:49:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM121523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 21 01:43:54 2023
Response via : Initial Calibration

