

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122121\
 Data File : VV024128.D
 Acq On : 21 Dec 2021 15:49
 Operator : SY/MD
 Sample : VSTD00240
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002240

Quant Time: Dec 22 03:20:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 22 03:17:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	8172	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	7889	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	4423	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	15503	1.851	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.019	65	11264	2.004	ug/L	-0.02
7) 1,2-Dichloropropane-d6	6.068	67	11074	1.695	ug/L	0.00
8) Toluene-d8	7.301	98	40496	2.243	ug/L	-0.02
10) 1,1,2,2-Tetrachloroeth...	10.199	84	7120	1.791	ug/L	-0.02
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	13628	1.537	ug/L	98
6) Trichloroethene	5.901	95	12319	1.897	ug/L	89
9) 1,2-Dibromoethane	8.353	107	7426	1.886	ug/L #	96
12) 1,2,3-Trichloropropane	10.267	75	5731	1.605	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.409	75	1054	1.623	ug/L #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122121\
Data File : VV024128.D
Acq On : 21 Dec 2021 15:49
Operator : SY/MD
Sample : VSTD00240
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002240

Quant Time: Dec 22 03:20:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 22 03:17:27 2021
Response via : Initial Calibration

