

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122121\
 Data File : VV024124.D
 Acq On : 21 Dec 2021 12:44
 Operator : SY/MD
 Sample : VSTD0.0536
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05236

Quant Time: Dec 22 03:19:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 22 03:17:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	9909	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	9672	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	5555	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	738	0.073	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	322	0.047	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	328	0.041	ug/L	0.00
8) Toluene-d8	7.317	98	1206	0.054	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	207	0.042	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.303	62	474	0.044	ug/L #	78
6) Trichloroethene	5.925	95	282	0.035	ug/L	89
9) 1,2-Dibromoethane	8.368	107	173	0.036	ug/L #	80
12) 1,2,3-Trichloropropane	10.267	75	136	0.030	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.409	75	30	0.037	ug/L #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122121\
Data File : VV024124.D
Acq On : 21 Dec 2021 12:44
Operator : SY/MD
Sample : VSTD0.0536
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05236

Quant Time: Dec 22 03:19:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 22 03:17:27 2021
Response via : Initial Calibration

