

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122121\
 Data File : VV024125.D
 Acq On : 21 Dec 2021 13:46
 Operator : SY/MD
 Sample : VSTD0.137
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1237

Quant Time: Dec 22 03:19:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 22 03:17:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	9532	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	9192	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	5220	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	933	0.096	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	646	0.099	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	647	0.085	ug/L	0.00
8) Toluene-d8	7.317	98	2264	0.108	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	402	0.087	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	866	0.084	ug/L	90
6) Trichloroethene	5.925	95	689	0.091	ug/L	94
9) 1,2-Dibromoethane	8.353	107	411	0.090	ug/L #	90
12) 1,2,3-Trichloropropane	10.267	75	355	0.084	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.426	75	73	0.095	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122121\
Data File : VV024125.D
Acq On : 21 Dec 2021 13:46
Operator : SY/MD
Sample : VSTD0.137
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1237

Quant Time: Dec 22 03:19:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 22 03:17:27 2021
Response via : Initial Calibration

