

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122121\
 Data File : VV024127.D
 Acq On : 21 Dec 2021 14:34
 Operator : SY/MD
 Sample : VSTD00139
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001239

Quant Time: Dec 22 03:19:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 22 03:17:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	8684	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	8337	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	4627	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	9280	1.043	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	7028	1.177	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	6797	0.985	ug/L	0.00
8) Toluene-d8	7.317	98	23841	1.249	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	4460	1.062	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	8260	0.877	ug/L	98
6) Trichloroethene	5.925	95	7193	1.048	ug/L	98
9) 1,2-Dibromoethane	8.353	107	4779	1.148	ug/L #	98
12) 1,2,3-Trichloropropane	10.267	75	3680	0.985	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.409	75	723	1.064	ug/L #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122121\
Data File : VV024127.D
Acq On : 21 Dec 2021 14:34
Operator : SY/MD
Sample : VSTD00139
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001239

Quant Time: Dec 22 03:19:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 22 03:17:27 2021
Response via : Initial Calibration

