

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122220\
 Data File : VV019791.D
 Acq On : 22 Dec 2020 12:51
 Operator : SY/MD
 Sample : L5173-04
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4N9

Quant Time: Dec 23 01:20:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 23 01:05:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	699639	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	655304	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	315885	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	253826	50.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.18%
7) Chloroethane-d5	1.57	69	203027	51.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.62%
11) 1,1-Dichloroethene-d2	2.11	63	366332	39.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.68%
21) 2-Butanone-d5	3.88	46	322251	118.18	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	118.18%
24) Chloroform-d	4.35	84	438387	48.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.96%
26) 1,2-Dichloroethane-d4	5.03	65	301946	51.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.20%
32) Benzene-d6	5.05	84	888063	50.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.62%
36) 1,2-Dichloropropane-d6	6.07	67	283699	51.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.64%
41) Toluene-d8	7.32	98	769196	48.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.76%
43) trans-1,3-Dichloropropene-	7.62	79	135626	43.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.10%
47) 2-Hexanone-d5	8.09	63	187059	87.06	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.06%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	313564	42.34	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.68%
64) 1,2-Dichlorobenzene-d4	11.63	152	277537	47.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.00%

Target Compounds					Ovalue
13) Acetone	2.18	43	6498	2.797 ug/L	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV122220\
Data File : VV019791.D
Acq On : 22 Dec 2020 12:51
Operator : SY/MD
Sample : L5173-04
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4N9

Quant Time: Dec 23 01:20:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 23 01:05:06 2020
Response via : Initial Calibration

