

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122220\
 Data File : VV019799.D
 Acq On : 22 Dec 2020 16:45
 Operator : SY/MD
 Sample : VIBLK76
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK76

Quant Time: Dec 23 02:13:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 23 01:05:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	723682	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	681239	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	322106	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	254316	49.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.02%
7) Chloroethane-d5	1.57	69	206799	50.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.06%
11) 1,1-Dichloroethene-d2	2.11	63	360633	37.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.88%
21) 2-Butanone-d5	3.89	46	375611	133.17	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	133.17%#
24) Chloroform-d	4.35	84	450942	48.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.44%
26) 1,2-Dichloroethane-d4	5.04	65	317357	51.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.84%
32) Benzene-d6	5.05	84	913849	50.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.60%
36) 1,2-Dichloropropane-d6	6.07	67	293956	51.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.30%
41) Toluene-d8	7.32	98	793329	48.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.00%
43) trans-1,3-Dichloropropene-	7.62	79	134198	41.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.90%
47) 2-Hexanone-d5	8.09	63	222671	99.68	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.68%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	348001	45.20	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.40%
64) 1,2-Dichlorobenzene-d4	11.63	152	288323	47.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.78%

Target Compounds					Ovalue
46) Tetrachloroethene	7.98	164	2841	0.781 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV122220\
Data File : VV019799.D
Acq On : 22 Dec 2020 16:45
Operator : SY/MD
Sample : VIBLK76
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK76

Quant Time: Dec 23 02:13:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 23 01:05:06 2020
Response via : Initial Calibration

