

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122221\
 Data File : VV024144.D
 Acq On : 22 Dec 2021 16:35
 Operator : SY/MD
 Sample : M5090-18
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3F1

Quant Time: Dec 23 06:14:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 23 06:11:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	10603	0.500	ug/L	0.02
5) Chlorobenzene-d5	8.864	117	10400	0.500	ug/L	0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	4981	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	4450	0.377	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 76.000%			
4) 1,2-Dichloroethane-d4	5.042	65	4106	0.541	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 108.000%			
7) 1,2-Dichloropropane-d6	6.068	67	4822	0.630	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 126.000%			
8) Toluene-d8	7.317	98	13527	0.496	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 100.000%			
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2726	0.559	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 112.000%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122221\
Data File : VV024144.D
Acq On : 22 Dec 2021 16:35
Operator : SY/MD
Sample : M5090-18
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3F1

Quant Time: Dec 23 06:14:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 23 06:11:19 2021
Response via : Initial Calibration

