

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122221\
 Data File : VV024150.D
 Acq On : 22 Dec 2021 18:58
 Operator : SY/MD
 Sample : M5152-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3E5

Quant Time: Dec 23 06:15:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 23 06:11:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	10478	0.500	ug/L	0.02
5) Chlorobenzene-d5	8.863	117	10158	0.500	ug/L	0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	5532	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	5109	0.438	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	88.000%
4) 1,2-Dichloroethane-d4	5.042	65	4173	0.556	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	112.000%
7) 1,2-Dichloropropane-d6	6.068	67	5041	0.674	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	134.000%
8) Toluene-d8	7.317	98	13936	0.523	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.000%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2857	0.600	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	120.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122221\
Data File : VV024150.D
Acq On : 22 Dec 2021 18:58
Operator : SY/MD
Sample : M5152-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3E5

Quant Time: Dec 23 06:15:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 23 06:11:19 2021
Response via : Initial Calibration

