

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122222\
 Data File : VV029757.D
 Acq On : 22 Dec 2022 11:42
 Operator : SY/MD
 Sample : VV1222WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK313

Quant Time: Dec 23 02:09:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 21 00:15:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	472454	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	446085	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	218727	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	124567	43.056	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	86.120%	
7) Chloroethane-d5	1.564	69	113857	51.611	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	103.220%	
11) 1,1-Dichloroethene-d2	2.105	63	171773	38.066	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	76.140%	
21) 2-Butanone-d5	3.876	46	189759	107.520	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	107.520%	
24) Chloroform-d	4.339	84	257049	47.214	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.420%	
26) 1,2-Dichloroethane-d4	5.024	65	155458	49.996	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.000%	
32) Benzene-d6	5.040	84	542667	46.039	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.080%	
36) 1,2-Dichloropropane-d6	6.059	67	178218	49.499	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	99.000%	
41) Toluene-d8	7.307	98	478524	43.087	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.180%	
43) trans-1,3-Dichloroprop...	7.612	79	72633	39.951	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	79.900%	
47) 2-Hexanone-d5	8.085	63	154192	100.290	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	100.290%	
56) 1,1,2,2-Tetrachloroeth...	10.204	84	255453	49.651	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.300%	
66) 1,2-Dichlorobenzene-d4	11.612	152	196113	47.100	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	94.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122222\
Data File : VV029757.D
Acq On : 22 Dec 2022 11:42
Operator : SY/MD
Sample : VV1222WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK313

Quant Time: Dec 23 02:09:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 21 00:15:13 2022
Response via : Initial Calibration

