

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122319\
 Data File : VV014176.D
 Acq On : 23 Dec 2019 22:43
 Operator : SY/MD
 Sample : K6404-16
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFE05

Quant Time: Dec 24 03:47:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 24 01:34:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	368276	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	370819	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	176173	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	101288	3.83	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.60%
7) Chloroethane-d5	1.58	69	90976	4.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.40%
11) 1,1-Dichloroethene-d2	2.13	63	131267	2.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.60%#
20) 2-Butanone-d5	3.93	46	196648	39.45	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.90%
24) Chloroform-d	4.40	84	194299	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.40%
26) 1,2-Dichloroethane-d4	5.08	65	96522	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.40%
32) Benzene-d6	5.10	84	398776	3.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.40%
36) 1,2-Dichloropropane-d6	6.11	67	120517	3.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.00%
41) Toluene-d8	7.36	98	337149	3.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.60%
43) trans-1,3-Dichloropropene-	7.66	79	37689	3.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	63.80%
46) 2-Hexanone-d5	8.13	63	151277	36.25	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.50%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	81814	3.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	139604	4.42	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.40%

Target Compounds					Ovalue
13) Acetone	2.19	43	19611	4.830 ug/L	98
14) Carbon disulfide	2.32	76	159154	2.255 ug/L	97
42) Toluene	7.43	91	12765	0.117 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV122319\
Data File : VV014176.D
Acq On : 23 Dec 2019 22:43
Operator : SY/MD
Sample : K6404-16
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFE05

Quant Time: Dec 24 03:47:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 24 01:34:37 2019
Response via : Initial Calibration

