

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122320\
 Data File : VV019816.D
 Acq On : 23 Dec 2020 16:57
 Operator : SY/MD
 Sample : L5206-01ME
 Misc : 4.53g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4L8ME

Quant Time: Dec 23 23:53:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 23 22:04:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	787348	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	744467	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	338976	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	247736	43.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.76%
7) Chloroethane-d5	1.57	69	165754	37.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.44%
11) 1,1-Dichloroethene-d2	2.08	63	360387	34.39	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.78%
21) 2-Butanone-d5	3.91	46	370425	120.71	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.71%
24) Chloroform-d	4.35	84	474175	46.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.04	65	340828	51.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.50%
32) Benzene-d6	5.05	84	966408	48.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.34%
36) 1,2-Dichloropropane-d6	6.07	67	312967	50.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.64%
41) Toluene-d8	7.32	98	844254	47.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.46%
43) trans-1,3-Dichloropropene-	7.63	79	143785	40.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.28%
47) 2-Hexanone-d5	8.13	63	246405	100.94	ug/L	0.04
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.94%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	364489	43.33	ug/L	0.01
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.66%
64) 1,2-Dichlorobenzene-d4	11.64	152	286022	45.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV122320\
Data File : VV019816.D
Acq On : 23 Dec 2020 16:57
Operator : SY/MD
Sample : L5206-01ME
Misc : 4.53µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4L8ME

Quant Time: Dec 23 23:53:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 23 22:04:07 2020
Response via : Initial Calibration

