

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122320\
 Data File : VV019817.D
 Acq On : 23 Dec 2020 17:22
 Operator : SY/MD
 Sample : VIBLK79
 Misc : 4.53g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK79

Quant Time: Dec 23 23:56:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 23 22:04:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	697429	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	652601	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	311011	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	227660	45.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.04%
7) Chloroethane-d5	1.57	69	191274	48.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.98%
11) 1,1-Dichloroethene-d2	2.11	63	331913	35.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.52%
21) 2-Butanone-d5	3.89	46	354543	130.43	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	130.43%#
24) Chloroform-d	4.35	84	433274	48.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.14%
26) 1,2-Dichloroethane-d4	5.04	65	304789	51.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.48%
32) Benzene-d6	5.05	84	868102	49.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.76%
36) 1,2-Dichloropropane-d6	6.07	67	277234	50.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.70%
41) Toluene-d8	7.32	98	748659	47.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.54%
43) trans-1,3-Dichloropropene-	7.62	79	125236	40.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.76%
47) 2-Hexanone-d5	8.09	63	207585	97.01	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.01%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	333821	45.27	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.54%
64) 1,2-Dichlorobenzene-d4	11.63	152	272641	46.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV122320\
Data File : VV019817.D
Acq On : 23 Dec 2020 17:22
Operator : SY/MD
Sample : VIBLK79
Misc : 4.53g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK79

Quant Time: Dec 23 23:56:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 23 22:04:07 2020
Response via : Initial Calibration

