

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122320\
 Data File : VV019815.D
 Acq On : 23 Dec 2020 16:33
 Operator : SY/MD
 Sample : L5206-06ME
 Misc : 5.25g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4R1ME

Quant Time: Dec 23 23:39:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 23 22:04:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	749579	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	705251	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	323047	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	247188	45.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.98%
7) Chloroethane-d5	1.57	69	175309	41.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.70%
11) 1,1-Dichloroethene-d2	2.08	63	361332	36.22	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.44%
21) 2-Butanone-d5	3.91	46	334277	114.42	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	114.42%
24) Chloroform-d	4.35	84	469127	48.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.86%
26) 1,2-Dichloroethane-d4	5.04	65	330050	52.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.26%
32) Benzene-d6	5.05	84	952890	50.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.32%
36) 1,2-Dichloropropane-d6	6.07	67	309124	52.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.92%
41) Toluene-d8	7.32	98	835645	49.34	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.68%
43) trans-1,3-Dichloropropene-	7.63	79	140109	41.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.62%
47) 2-Hexanone-d5	8.14	63	222403	96.17	ug/L	0.05
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.17%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	350273	43.95	ug/L	0.01
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.90%
64) 1,2-Dichlorobenzene-d4	11.64	152	284059	47.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.08%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV122320\
Data File : VV019815.D
Acq On : 23 Dec 2020 16:33
Operator : SY/MD
Sample : L5206-06ME
Misc : 5.25g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4R1ME

Quant Time: Dec 23 23:39:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 23 22:04:07 2020
Response via : Initial Calibration

