

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\
 Method File : SFAMVSIM122421.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Dec 24 15:33:30 2021
 Response Via : Initial Calibration

Calibration Files

0.5 =VV024156.D 0.05=VV024154.D 0.1 =VV024155.D 1.0 =VV024157.D 2 =VV024158.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.564	0.636	0.719	0.696	0.499	0.623	14.70
3) T	Vinyl chloride	0.600	0.682	0.602	0.664	0.487	0.607	12.59
4) S	1,2-Dichloroetha...	0.435	0.425	0.450	0.366	0.444	0.424	7.96
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.488	0.483	0.395	0.368	0.488	0.444	13.10
7) S	1,2-Dichloroprop...	0.503	0.493	0.547	0.367	0.501	0.482	14.05
8) S	Toluene-d8	1.342	1.408	1.439	1.205	1.405	1.360	6.87
9) T	1,2-Dibromoethane	0.320	0.318	0.253	0.245	0.316	0.290	13.04
10) S	1,1,2,2-Tetrachl...	0.317	0.319	0.335	0.248	0.309	0.306	10.98
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.513	0.473	0.397	0.353	0.508	0.449	15.81
13) T	1,2-Dibromo-3-ch...	0.089	0.096	0.076	0.066	0.086	0.083	14.19

(#) = Out of Range