

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122421\
 Data File : VV024158.D
 Acq On : 24 Dec 2021 14:38
 Operator : SY/MD
 Sample : VSTD00246
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001246

Quant Time: Dec 24 15:29:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 24 15:27:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	11531	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	11500	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	6040	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	23022	1.603	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	20494	2.095	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	23039	2.077	ug/L	0.00
8) Toluene-d8	7.317	98	64645	2.067	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	14209	2.022	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	22462	1.605	ug/L	97
6) Trichloroethene	5.901	95	22450	2.196	ug/L	82
9) 1,2-Dibromoethane	8.353	107	14541	2.176	ug/L #	97
12) 1,2,3-Trichloropropane	10.267	75	12284	2.267	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.409	75	2083	2.089	ug/L #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122421\
Data File : VV024158.D
Acq On : 24 Dec 2021 14:38
Operator : SY/MD
Sample : VSTD00246
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001246

Quant Time: Dec 24 15:29:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 24 15:27:15 2021
Response via : Initial Calibration

