

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122421\
 Data File : VV024165.D
 Acq On : 24 Dec 2021 17:48
 Operator : SY/MD
 Sample : M5152-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3C1

Quant Time: Dec 24 21:47:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 24 15:33:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	10348	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	9945	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	5459	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	5028	0.390	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	78.000%
4) 1,2-Dichloroethane-d4	5.043	65	3920	0.447	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	90.000%
7) 1,2-Dichloropropane-d6	6.069	67	4587	0.478	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	96.000%
8) Toluene-d8	7.317	98	12448	0.460	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2922	0.481	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	96.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	16766	1.335	ug/L	94
6) Trichloroethene	5.925	95	2955	0.334	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122421\
Data File : VV024165.D
Acq On : 24 Dec 2021 17:48
Operator : SY/MD
Sample : M5152-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3C1

Quant Time: Dec 24 21:47:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 24 15:33:30 2021
Response via : Initial Calibration

