

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122421\
 Data File : VV024166.D
 Acq On : 24 Dec 2021 18:12
 Operator : SY/MD
 Sample : M5152-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3C4

Quant Time: Dec 24 21:47:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 24 15:33:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	10501	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	10000	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	5456	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	5155	0.394	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	78.000%		
4) 1,2-Dichloroethane-d4	5.042	65	3835	0.431	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	86.000%		
7) 1,2-Dichloropropane-d6	6.069	67	4590	0.476	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	96.000%		
8) Toluene-d8	7.317	98	12250	0.450	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	90.000%		
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2874	0.470	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	94.000%		
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	24111	1.892	ug/L	95
6) Trichloroethene	5.925	95	2892	0.325	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122421\
Data File : VV024166.D
Acq On : 24 Dec 2021 18:12
Operator : SY/MD
Sample : M5152-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3C4

Quant Time: Dec 24 21:47:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 24 15:33:30 2021
Response via : Initial Calibration

