

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122421\
 Data File : VV024168.D
 Acq On : 24 Dec 2021 19:00
 Operator : SY/MD
 Sample : M5152-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3D4

Quant Time: Dec 24 21:47:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 24 15:33:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	10511	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	10132	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	5520	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	5425	0.414	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	82.000%		
4) 1,2-Dichloroethane-d4	5.042	65	3310	0.371	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	74.000%		
7) 1,2-Dichloropropane-d6	6.068	67	4729	0.484	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	96.000%		
8) Toluene-d8	7.317	98	12678	0.460	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	92.000%		
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2981	0.481	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	96.000%		
Target Compounds						
3) Vinyl chloride	1.310	62	61179	4.796	ug/L	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122421\
Data File : VV024168.D
Acq On : 24 Dec 2021 19:00
Operator : SY/MD
Sample : M5152-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3D4

Quant Time: Dec 24 21:47:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 24 15:33:30 2021
Response via : Initial Calibration

