

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122421\
 Data File : VV024156.D
 Acq On : 24 Dec 2021 13:51
 Operator : SY/MD
 Sample : VSTD0.544
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5244

Quant Time: Dec 24 15:29:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 24 15:27:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	11576	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	11403	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	6054	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	6529	0.453	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	5031	0.512	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	5741	0.522	ug/L	0.00
8) Toluene-d8	7.317	98	15304	0.493	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	3619	0.519	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	6942	0.494	ug/L	98
6) Trichloroethene	5.925	95	5569	0.549	ug/L	96
9) 1,2-Dibromoethane	8.353	107	3649	0.551	ug/L #	98
12) 1,2,3-Trichloropropane	10.267	75	3103	0.571	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.426	75	537	0.537	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122421\
Data File : VV024156.D
Acq On : 24 Dec 2021 13:51
Operator : SY/MD
Sample : VSTD0.544
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5244

Quant Time: Dec 24 15:29:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM122421.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 24 15:27:15 2021
Response via : Initial Calibration

