

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122721\
 Data File : VW024190.D
 Acq On : 27 Dec 2021 19:26
 Operator : SY/MD
 Sample : M5170-03DL 100X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AN8DL

Quant Time: Dec 28 05:52:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 28 05:23:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	124253	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	128044	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	65597	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	38102	4.077	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.600%
7) Chloroethane-d5	1.568	69	35025	4.494	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.800%
11) 1,1-Dichloroethene-d2	2.108	63	58561	2.578	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.600%#
20) 2-Butanone-d5	3.905	46	64125	61.769	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.540%
24) Chloroform-d	4.349	84	90920	4.583	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.034	65	44721	5.036	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.050	84	171219	4.952	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	6.069	67	49338	5.209	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.200%
41) Toluene-d8	7.317	98	152455	4.473	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	7.625	79	18139	4.523	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.400%
46) 2-Hexanone-d5	8.092	63	72675	53.892	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.780%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	34172	5.284	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	60733	5.326	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.600%
Target Compounds						
18) trans-1,2-Dichloroethene	2.764	96	3059	0.169	ug/L	89
22) cis-1,2-Dichloroethene	3.908	96	534436	29.807	ug/L	100
34) Trichloroethene	5.915	95	13527	0.714	ug/L	93
47) Tetrachloroethene	7.976	164	16269	0.965	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122721\
Data File : VV024190.D
Acq On : 27 Dec 2021 19:26
Operator : SY/MD
Sample : M5170-03DL 100X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AN8DL

Quant Time: Dec 28 05:52:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 28 05:23:55 2021
Response via : Initial Calibration

