

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122721\
 Data File : VW024191.D
 Acq On : 27 Dec 2021 19:50
 Operator : SY/MD
 Sample : M5170-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Dec 28 05:52:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 28 05:23:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	123398	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	124768	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	64095	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	37818	4.075	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.400%
7) Chloroethane-d5	1.568	69	34923	4.512	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.200%
11) 1,1-Dichloroethene-d2	2.108	63	58626	2.599	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.000%#
20) 2-Butanone-d5	3.921	46	53664	52.051	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.100%
24) Chloroform-d	0.000	84	0d	0.000	ug/L	
Spiked Amount	5.000	Range	70 - 125	Recovery	=	0.000%#
26) 1,2-Dichloroethane-d4	5.034	65	43865	4.973	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%
32) Benzene-d6	5.050	84	168221	4.993	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
36) 1,2-Dichloropropane-d6	6.069	67	48539	5.259	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%
41) Toluene-d8	7.317	98	150699	4.538	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	7.625	79	17173	4.394	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.800%
46) 2-Hexanone-d5	8.091	63	70446	53.611	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.220%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	33209	5.270	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	61080	5.482	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122721\
Data File : VV024191.D
Acq On : 27 Dec 2021 19:50
Operator : SY/MD
Sample : M5170-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Dec 28 05:52:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 28 05:23:55 2021
Response via : Initial Calibration

