

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122819\
 Data File : VV014210.D
 Acq On : 28 Dec 2019 12:07
 Operator : SY/MD
 Sample : VV1228WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK41

Quant Time: Dec 30 09:00:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Dec 30 00:30:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	852273	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	781709	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	360891	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	179987	4.51	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.20%
7) Chloroethane-d5	1.58	69	161817	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.40%
11) 1,1-Dichloroethene-d2	2.13	63	253707	3.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.80%
20) 2-Butanone-d5	3.95	46	436032	43.98	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.96%
24) Chloroform-d	4.40	84	445620	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
26) 1,2-Dichloroethane-d4	5.08	65	233838	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
32) Benzene-d6	5.09	84	989699	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
36) 1,2-Dichloropropane-d6	6.11	67	307022	4.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.00%
41) Toluene-d8	7.35	98	919438	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
43) trans-1,3-Dichloropropene-	7.66	79	132083	4.57	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.40%
46) 2-Hexanone-d5	8.13	63	506164	46.14	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.28%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	213130	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	303200	4.69	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.80%

Target Compounds					Ovalue
16) Methylene chloride	2.53	84	17123	0.270 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122819\
Data File : VV014210.D
Acq On : 28 Dec 2019 12:07
Operator : SY/MD
Sample : VV1228WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK41

Quant Time: Dec 30 09:00:10 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Dec 30 00:30:20 2019
Response via : Initial Calibration

