

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122820\
 Data File : VV019850.D
 Acq On : 28 Dec 2020 15:12
 Operator : SY/MD
 Sample : VV1228WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK66

Quant Time: Dec 29 03:18:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 29 03:03:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	620632	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	584761	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	280700	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	236376	49.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.44%
7) Chloroethane-d5	1.57	69	192180	50.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.68%
11) 1,1-Dichloroethene-d2	2.11	63	337412	37.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.34%
21) 2-Butanone-d5	3.89	46	364957	102.52	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.52%
24) Chloroform-d	4.35	84	427757	47.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.28%
26) 1,2-Dichloroethane-d4	5.04	65	301168	50.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.56%
32) Benzene-d6	5.05	84	857044	50.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.07	67	276483	49.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.86%
41) Toluene-d8	7.32	98	748405	49.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.88%
43) trans-1,3-Dichloropropene-	7.62	79	132360	48.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.18%
47) 2-Hexanone-d5	8.09	63	212475	92.85	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.85%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	337522	46.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.74%
64) 1,2-Dichlorobenzene-d4	11.63	152	270648	50.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV122820\
Data File : VV019850.D
Acq On : 28 Dec 2020 15:12
Operator : SY/MD
Sample : VV1228WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK66

Quant Time: Dec 29 03:18:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 29 03:03:45 2020
Response via : Initial Calibration

