

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122821\
 Data File : VW024226.D
 Acq On : 28 Dec 2021 18:22
 Operator : SY/MD
 Sample : M5193-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0J13

Quant Time: Dec 29 05:07:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 29 04:43:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	82575	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	84691	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	45510	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	24745	3.984	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.600%
7) Chloroethane-d5	1.564	69	23844	4.603	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.104	63	53617	3.552	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.000%
20) 2-Butanone-d5	3.902	46	41826	60.625	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.240%
24) Chloroform-d	4.346	84	62161	4.715	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.200%
26) 1,2-Dichloroethane-d4	5.034	65	29028	4.918	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
32) Benzene-d6	5.050	84	109964	4.808	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%
36) 1,2-Dichloropropane-d6	6.069	67	31440	5.019	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.400%
41) Toluene-d8	7.316	98	97835	4.340	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	7.625	79	10219	3.852	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.000%
46) 2-Hexanone-d5	8.091	63	45026	50.481	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.960%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	23152	5.413	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	41753	5.278	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.600%
Target Compounds						Qvalue
5) Vinyl chloride	1.307	62	34788	2.357	ug/L	# 23
12) 1,1-Dichloroethene	2.114	96	28603	2.523	ug/L	95
13) Acetone	2.204	43	15391	15.719	ug/L	76
14) Carbon disulfide	2.291	76	26503	0.723	ug/L	96
17) Methyl tert-butyl Ether	2.767	73	1525	0.067	ug/L	# 21
18) trans-1,2-Dichloroethene	2.757	96	204715	17.001	ug/L	97
19) 1,1-Dichloroethane	3.185	63	102082	5.093	ug/L	99
21) 2-Butanone	4.005	43	15637	12.138	ug/L	86
22) cis-1,2-Dichloroethene	3.905	96	1985836	166.655	ug/L	100
34) Trichloroethene	5.908	95	381669	30.443	ug/L	98
42) Toluene	7.394	91	5294	0.107	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122821\
Data File : VV024226.D
Acq On : 28 Dec 2021 18:22
Operator : SY/MD
Sample : M5193-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0J13

Quant Time: Dec 29 05:07:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 29 04:43:10 2021
Response via : Initial Calibration

