

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122821\
 Data File : VW024229.D
 Acq On : 28 Dec 2021 19:34
 Operator : SY/MD
 Sample : VIBLK
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK249

Quant Time: Dec 29 05:07:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 29 04:43:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	82662	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	87557	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	46028	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	25756	4.142	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	82.800%	
7) Chloroethane-d5	1.568	69	24449	4.715	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.400%	
11) 1,1-Dichloroethene-d2	2.108	63	41182	2.725	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	54.600%#	
20) 2-Butanone-d5	3.902	46	59557	86.234	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	172.460%#	
24) Chloroform-d	4.349	84	63351	4.800	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.000%	
26) 1,2-Dichloroethane-d4	5.034	65	31780	5.379	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.600%	
32) Benzene-d6	5.050	84	111890	4.732	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.600%	
36) 1,2-Dichloropropane-d6	6.069	67	32993	5.094	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.800%	
41) Toluene-d8	7.317	98	97499	4.184	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
43) trans-1,3-Dichloroprop...	7.625	79	9712	3.541	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	70.800%	
46) 2-Hexanone-d5	8.091	63	46557	50.488	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	100.980%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	22938	5.187	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.800%	
66) 1,2-Dichlorobenzene-d4	11.625	152	40243	5.030	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122821\
Data File : VV024229.D
Acq On : 28 Dec 2021 19:34
Operator : SY/MD
Sample : VIBLK
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK249

Quant Time: Dec 29 05:07:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 29 04:43:10 2021
Response via : Initial Calibration

