

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122821\
 Data File : VW024230.D
 Acq On : 28 Dec 2021 19:58
 Operator : SY/MD
 Sample : M5233-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Dec 29 05:07:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 29 04:43:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	78572	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	80540	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	41541	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	26017	4.402	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.000%
7) Chloroethane-d5	1.568	69	25813	5.237	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.800%
11) 1,1-Dichloroethene-d2	2.108	63	43102	3.001	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.000%
20) 2-Butanone-d5	3.902	46	60567	92.261	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	184.520%#
24) Chloroform-d	4.349	84	65455	5.217	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.400%
26) 1,2-Dichloroethane-d4	5.034	65	31945	5.688	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.800%
32) Benzene-d6	5.050	84	117123	5.385	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.600%
36) 1,2-Dichloropropane-d6	6.069	67	33074	5.551	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.000%
41) Toluene-d8	7.317	98	101990	4.758	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
43) trans-1,3-Dichloroprop...	7.625	79	10409	4.126	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.600%
46) 2-Hexanone-d5	8.088	63	49173	57.971	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.940%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	24348	5.986	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	119.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	42299	5.858	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122821\
Data File : VV024230.D
Acq On : 28 Dec 2021 19:58
Operator : SY/MD
Sample : M5233-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Dec 29 05:07:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 29 04:43:10 2021
Response via : Initial Calibration

